



US 20170100101A1

(19) **United States**

(12) **Patent Application Publication**

RYOO et al.

(10) **Pub. No.: US 2017/0100101 A1**

(43) **Pub. Date: Apr. 13, 2017**

(54) **ULTRASOUND DIAGNOSIS METHOD AND APPARATUS FOR ANALYZING CONTRAST ENHANCED ULTRASOUND IMAGE**

(71) Applicant: **SAMSUNG MEDISON CO., LTD.**, Hongcheon-gun (KR)

(72) Inventors: **Ho-jin RYOO**, Hongcheon-gun (KR); **Jong-sik KIM**, Hongcheon-gun (KR); **Jin-yong LEE**, Hongcheon-gun (KR)

(73) Assignee: **SAMSUNG MEDISON CO., LTD.**, Hongcheon-gun (KR)

(21) Appl. No.: **15/082,174**

(22) Filed: **Mar. 28, 2016**

(30) **Foreign Application Priority Data**

Oct. 8, 2015 (KR) 10-2015-0141637

Publication Classification

(51) **Int. Cl.**

A61B 8/08 (2006.01)

A61B 8/14 (2006.01)

A61B 8/00 (2006.01)

(52) **U.S. Cl.**

CPC *A61B 8/5215* (2013.01); *A61B 8/481* (2013.01); *A61B 8/469* (2013.01); *A61B 8/465* (2013.01); *A61B 8/5207* (2013.01); *A61B 8/14* (2013.01); *A61B 8/085* (2013.01)

(57) **ABSTRACT**

Provided are an ultrasound diagnosis apparatus and method for analyzing a contrast enhanced ultrasound image. The ultrasound diagnosis apparatus and method determine, by analysis, causes of defects in a frame of an ultrasound image based on a time intensity curve and allow visual display of the defective frame based on the determined causes. Furthermore, the ultrasound diagnosis method and apparatus may provide a user interface for quick access to causes of defects in a frame of the ultrasound image and to a position of the defective frame.

1000

